



Specification for Model 2003™ - 2006™:

The Platform Scale must have rugged steel frame and shall be dry powder polyester plastic coated. The decking shall be solid, 1/2 inch thick PVC. All hardware shall be stainless steel. The height shall not exceed 1.75 inches, (45 mm) so as not to require a pit or ramp. Stainless steel leveling feet shall be supplied with 4 NTEP approved stainless steel shear beam cells. A four cell system is required for accurate reading so that off center loads may be placed anywhere on the base and still obtain an accurate reading. Base must be fully electronic with no mechanical lever system to corrode.

Scale shall be a Model 2003-2006™ as manufactured by Scaleton Industries Ltd., Plumsteadville, PA.