



Specification for Model 2305™:

Scale must weigh 150 lb. cylinder. The accuracy must be at least 0.5%. Digital readout must be an LCD Display with characters that are 0.5 inches high, with 0.1 lb. or 0.1 kg resolution. Indicator is required to have 4-20 mA output selectable as scale powered or loop powered, standard. Indicator must have a knob to set tare weight. The electronics must be housed in a NEMA 4X, UL approved enclosure. All base hardware must be stainless steel. The base must be dry powder polyester plastic coated steel with PVC decking to prevent corrosion.

Scale shall be a Model 2305™ as manufactured by Scaleton Industries Ltd., Plumsteadville, PA.