



### **Specification for Model 4000SS™:**

The Gas Cabinet, Dewar & Cylinder Scale must be constructed of all heavy gauge stainless steel. All hardware shall be stainless steel to prevent corrosion. The profile shall not exceed 1.75 inches (44.5 mm) so as not to require a pit or ramp. The load cell must be IP65 approved. Base must be fully electronic with no mechanical lever system to corrode.

Scale shall be a Model 4000SS™ as manufactured by Scaleton Industries Ltd., Plumsteadville, PA.

**Scaleton™ Industries, Ltd.**  
Bedminster Industrial Park  
53 Apple Tree Lane  
P.O. Box 365  
Plumsteadville, PA 18949 USA

**Toll Free: 1-800-257-5911 (USA & Canada)**  
Phone: (+1) 215-766-2670 (international)  
Fax: (+1) 215-766-2672  
E-mail: [info@scaletonscales.com](mailto:info@scaletonscales.com)  
Website: [www.scaletonscales.com](http://www.scaletonscales.com)