



Specification for Model 1020-B™:

Model 1020-B™ V5, 5 Digit LCD Controller must be electronic with a backlit, two line, 16 character, alphanumeric digital display, housed in a NEMA 4X, UL approved enclosure. Indicator shall have three methods of dispensing, with one or two relays to provide control for fast and/or slow fill process. Controls shall be provided by a 6 button keypad and will have four LED lights to indicate current process. 4-20mA output shall be standard and reflect the net weight of product in container. Accuracy is 0.1% to 0.25% of full scale capacity. Indicator will be factory calibrated and programmed, and will have factory calibration restore feature that does not require test weights.

Indicator will be the 5 Digit Controller Model 1020-B™ V5 as manufactured by Scaleton Industries Ltd., Plumsteadville, PA.