



### **Specification for Model 1020™:**

Model 1020™, 5 Digit LCD Controller must be electronic with a two line, backlit, alphanumeric digital display, housed in a NEMA 4x, UL approved enclosure. Available as single or dual channel and capable of displaying two channels simultaneously. A loop powered, programmable output of 4-20mA and 4 HI or LO set points available. Indicator must be capable of 0.1 lb, kg, gal., or ltr. Accuracy is 0.1% of full scale capacity. Indicator can be calibrated in the field with no test weights. Indicator must be CSA certified.

Indicator shall be a Model 1020™ as manufactured by Scaleton Industries Ltd., Plumsteadville, PA.